

GC Fuji PLUS™ EWT

RADIOPAQUE REINFORCED GLASS IONOMER LUTING CEMENT WITH EXTENDED WORKING TIME

Use only by a dental professional in the indications for use.

INDICATIONS FOR USE

- Cementation of metal-based inlays, onlays, crowns and bridges
- Cementation of resin inlays, onlays, crowns and bridges
- Cementation of all ceramic inlays
- Cementation of high strength (zirconia based) all ceramic crowns and bridges
- Cementation of metal, ceramic and fibre based

CONTRAINDICATIONS

- Direct pulp capping
- Avoid use of this product in patients with known allergies to glass ionomer cement, methacrylic monomer or methacrylic polymer.

COMPOSITION
Powder: Fluoro-alumino-silicate glass, initiator, pigment
Liquid: Methacrylate, water, polyacrylic acid, dimethacrylate, carboxylic acid, stabilizer.

DIRECTIONS FOR USE

Powder / Liquid Ratio (g/g)	2.0/1.0
Mixing Time (sec.)	30*
Working time (min., sec.) (at 23 °C) (from start of mixing)	20*
Setting Time (37°C/99°F) (min., sec.)	5'45"

Radio-opaque: Its radiopacity is equivalent to 2.1 mm or 210% aluminum. (The radiopacity of aluminum is comparable to dentin.)
ISO 9917-2 Class 1, Luting

- TOOTH PREPARATION**
a) Prepare tooth in usual manner. For pulp capping, use calcium hydroxide.

- clean the prepared teeth with pumice and water.

- Use GC Fuji PLUS CONDITIONER on the prepared tooth surface for 20 seconds for optimal adhesion.
- Rinse thoroughly with water. Remove excess moisture by blotting with a cotton pellet or gently blowing with an oil-free air. DO NOT RESCATO. Prepared surface should appear moist (glistening).

RESTORATION PREPARATION

Make sure that the restoration is pretreated and handled according to the manufacturer's instructions.

POWDER AND LIQUID DISPENSING

Note: Use GC Fuji PLUS Liquid. Refer to the instructions for use of GC Fuji PLUS.

- The standard powder to liquid ratio is 2.0g / 1.0g (1 level large scoop of powder to 3 drops of liquid, or 1 level small scoop of powder to 1.5 drops of liquid).
- For accurate dispensing of powder, lightly tap the bottle against the hand. Do not shake or invert.
- Use the liquid bottle vertically and squeeze gently.
- Close bottles immediately after use.

- Dispense powder and liquid onto the pad. Using the plastic spatula, add all the powder to the liquid. Mix rapidly for 20 seconds. Note: When using larger amounts, divide the powder into 2 equal parts. Mix the first portion with all of the liquid for 15 seconds. Incorporate the remaining portion and mix the whole thoroughly for 15-25 seconds (total 30-40 seconds).

- CEMENTATION**
a) Coat the internal surface of the restoration with sufficient cement and seal immediately. The working time is 3 minutes 30 seconds after start of mixing at 23°C (73.4°F). Higher temperatures will shorten working time.
b) Use moderate pressure.
- Start removing excess cement when the excess cement feels rubbery.
- Finishing can be started 5 minutes 45 seconds after sealing the restoration.

STORAGE
For optimum or optimal performance, store in a cool and dark place (25°C) (39.2-77.0°F).

SHADES

A3 (Vita® shade)
Vita® is a registered trademark of Vita Zahnfabrik, Bad Säckingen, Germany.

PACKAGE
Powder 15g with scoop

CAUTION

- In case of contact with oral tissue or skin, remove immediately with a sponge or cotton swab in alcohol. Flush with water. To avoid contact, a rubber dam and/or cocoa butter can be used to isolate the operation field from the skin or oral tissue.
- In case of contact with eyes, flush immediately with water and seek medical attention.
- DO NOT mix powder or liquid with components of other glass ionomer cements.

- GC Fuji PLUS CONDITIONER contains 2% Ferric Chloride. It is not recommended to use GC Fuji PLUS CONDITIONER in combination with products containing Epinephrine as this may stain oral tissue.
- Do not use GC Fuji PLUS EWT in combination with eugenol containing materials as eugenol may hinder GC Fuji PLUS EWT from setting or bonding properly.

- This product is not indicated for filling or core build-up.
- In case it stains the product may cause sensitivity in some people. If any such reactions are experienced, discontinue the use of the product and refer to a physician.

- Personal protective equipment (PPE) such as gloves, face masks and safety eyewear should always be worn.

Some products referenced in the present IFU may be classified as hazard-ous according to GHS. Always familiarize yourself with the safety data sheets available at:
http://www.goregore.com

They can also be obtained from your supplier.

CLEANING AND DISINFECTING.
MULTI-USE DELIVERY SYSTEMS: To avoid cross-contamination between patients this device requires mid-level disinfection. Immediately after use inspect device and label for deterioration. Discard device if damaged. DO NOT IMMERSÉ. Thoroughly clean device to prevent drying and accumulation of contaminants. Disinfect with a mid-level registered healthcare-grade infection control product according to regional/national guidelines.

Some products referenced in the present IFU may be classified as hazard-ous according to GHS. Always familiarize yourself with the safety data sheets available at:
http://www.goregore.com
They can also be obtained from your supplier.

For the Summary of Safety and Clinical Performance (SSCP) please see EUDAMED database (https://ec.europa.eu/tools/eudamed/) or contact us at Regulatory.gcc@gc.dental

Undesired effects - Reporting:
If you become aware of any kind of undesired effect, reaction or similar events experienced by use of this product, including those not listed in this instruction for use, please report them directly through the relevant vigilance system, by selecting the proper authority of your country according to the following link:

https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
In this way you will contribute to improve the safety of this product.

Last revised: 12/2023

For the Summary of Safety and Clinical Performance (SSCP) please see EUDAMED database (https://ec.europa.eu/tools/eudamed/) or contact us at Regulatory.gcc@gc.dental

Undesired effects - Reporting:
If you become aware of any kind of undesired effect, reaction or similar events experienced by use of this product, including those not listed in this instruction for use, please report them directly through the relevant vigilance system, by selecting the proper authority of your country according to the following link:

https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
In this way you will contribute to improve the safety of this product.

Last revised: 12/2023

For the Summary of Safety and Clinical Performance (SSCP) please see EUDAMED database (https://ec.europa.eu/tools/eudamed/) or contact us at Regulatory.gcc@gc.dental

Undesired effects - Reporting:
If you become aware of any kind of undesired effect, reaction or similar events experienced by use of this product, including those not listed in this instruction for use, please report them directly through the relevant vigilance system, by selecting the proper authority of your country according to the following link:

https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
In this way you will contribute to improve the safety of this product.

Last revised: 12/2023

For the Summary of Safety and Clinical Performance (SSCP) please see EUDAMED database (https://ec.europa.eu/tools/eudamed/) or contact us at Regulatory.gcc@gc.dental

Undesired effects - Reporting:
If you become aware of any kind of undesired effect, reaction or similar events experienced by use of this product, including those not listed in this instruction for use, please report them directly through the relevant vigilance system, by selecting the proper authority of your country according to the following link:

https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
In this way you will contribute to improve the safety of this product.

Last revised: 12/2023

For the Summary of Safety and Clinical Performance (SSCP) please see EUDAMED database (https://ec.europa.eu/tools/eudamed/) or contact us at Regulatory.gcc@gc.dental

Undesired effects - Reporting:
If you become aware of any kind of undesired effect, reaction or similar events experienced by use of this product, including those not listed in this instruction for use, please report them directly through the relevant vigilance system, by selecting the proper authority of your country according to the following link:

https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
In this way you will contribute to improve the safety of this product.

Last revised: 12/2023

For the Summary of Safety and Clinical Performance (SSCP) please see EUDAMED database (https://ec.europa.eu/tools/eudamed/) or contact us at Regulatory.gcc@gc.dental

Undesired effects - Reporting:
If you become aware of any kind of undesired effect, reaction or similar events experienced by use of this product, including those not listed in this instruction for use, please report them directly through the relevant vigilance system, by selecting the proper authority of your country according to the following link:

https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
In this way you will contribute to improve the safety of this product.

Last revised: 12/2023

For the Summary of Safety and Clinical Performance (SSCP) please see EUDAMED database (https://ec.europa.eu/tools/eudamed/) or contact us at Regulatory.gcc@gc.dental

Undesired effects - Reporting:
If you become aware of any kind of undesired effect, reaction or similar events experienced by use of this product, including those not listed in this instruction for use, please report them directly through the relevant vigilance system, by selecting the proper authority of your country according to the following link:

https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
In this way you will contribute to improve the safety of this product.

Last revised: 12/2023

For the Summary of Safety and Clinical Performance (SSCP) please see EUDAMED database (https://ec.europa.eu/tools/eudamed/) or contact us at Regulatory.gcc@gc.dental

Undesired effects - Reporting:
If you become aware of any kind of undesired effect, reaction or similar events experienced by use of this product, including those not listed in this instruction for use, please report them directly through the relevant vigilance system, by selecting the proper authority of your country according to the following link:

https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
In this way you will contribute to improve the safety of this product.

Last revised: 12/2023

For the Summary of Safety and Clinical Performance (SSCP) please see EUDAMED database (https://ec.europa.eu/tools/eudamed/) or contact us at Regulatory.gcc@gc.dental

Undesired effects - Reporting:
If you become aware of any kind of undesired effect, reaction or similar events experienced by use of this product, including those not listed in this instruction for use, please report them directly through the relevant vigilance system, by selecting the proper authority of your country according to the following link:

https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
In this way you will contribute to improve the safety of this product.

Last revised: 12/2023

For the Summary of Safety and Clinical Performance (SSCP) please see EUDAMED database (https://ec.europa.eu/tools/eudamed/) or contact us at Regulatory.gcc@gc.dental

Undesired effects - Reporting:
If you become aware of any kind of undesired effect, reaction or similar events experienced by use of this product, including those not listed in this instruction for use, please report them directly through the relevant vigilance system, by selecting the proper authority of your country according to the following link:

https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
In this way you will contribute to improve the safety of this product.

Last revised: 12/2023

For the Summary of Safety and Clinical Performance (SSCP) please see EUDAMED database (https://ec.europa.eu/tools/eudamed/) or contact us at Regulatory.gcc@gc.dental

Undesired effects - Reporting:
If you become aware of any kind of undesired effect, reaction or similar events experienced by use of this product, including those not listed in this instruction for use, please report them directly through the relevant vigilance system, by selecting the proper authority of your country according to the following link:

https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
In this way you will contribute to improve the safety of this product.

Last revised: 12/2023

For the Summary of Safety and Clinical Performance (SSCP) please see EUDAMED database (https://ec.europa.eu/tools/eudamed/) or contact us at Regulatory.gcc@gc.dental

Undesired effects - Reporting:
If you become aware of any kind of undesired effect, reaction or similar events experienced by use of this product, including those not listed in this instruction for use, please report them directly through the relevant vigilance system, by selecting the proper authority of your country according to the following link:

https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
In this way you will contribute to improve the safety of this product.

Last revised: 12/2023

For the Summary of Safety and Clinical Performance (SSCP) please see EUDAMED database (https://ec.europa.eu/tools/eudamed/) or contact us at Regulatory.gcc@gc.dental

Undesired effects - Reporting:
If you become aware of any kind of undesired effect, reaction or similar events experienced by use of this product, including those not listed in this instruction for use, please report them directly through the relevant vigilance system, by selecting the proper authority of your country according to the following link:

https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
In this way you will contribute to improve the safety of this product.

Last revised: 12/2023

For the Summary of Safety and Clinical Performance (SSCP) please see EUDAMED database (https://ec.europa.eu/tools/eudamed/) or contact us at Regulatory.gcc@gc.dental

Undesired effects - Reporting:
If you become aware of any kind of undesired effect, reaction or similar events experienced by use of this product, including those not listed in this instruction for use, please report them directly through the relevant vigilance system, by selecting the proper authority of your country according to the following link:

https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
In this way you will contribute to improve the safety of this product.

Last revised: 12/2023

For the Summary of Safety and Clinical Performance (SSCP) please see EUDAMED database (https://ec.europa.eu/tools/eudamed/) or contact us at Regulatory.gcc@gc.dental

Undesired effects - Reporting:
If you become aware of any kind of undesired effect, reaction or similar events experienced by use of this product, including those not listed in this instruction for use, please report them directly through the relevant vigilance system, by selecting the proper authority of your country according to the following link:

https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
In this way you will contribute to improve the safety of this product.

Last revised: 12/2023

For the Summary of Safety and Clinical Performance (SSCP) please see EUDAMED database (https://ec.europa.eu/tools/eudamed/) or contact us at Regulatory.gcc@gc.dental

Undesired effects - Reporting:
If you become aware of any kind of undesired effect, reaction or similar events experienced by use of this product, including those not listed in this instruction for use, please report them directly through the relevant vigilance system, by selecting the proper authority of your country according to the following link:

https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
In this way you will contribute to improve the safety of this product.

Last revised: 12/2023

For the Summary of Safety and Clinical Performance (SSCP) please see EUDAMED database (https://ec.europa.eu/tools/eudamed/) or contact us at Regulatory.gcc@gc.dental

Undesired effects - Reporting:
If you become aware of any kind of undesired effect, reaction or similar events experienced by use of this product, including those not listed in this instruction for use, please report them directly through the relevant vigilance system, by selecting the proper authority of your country according to the following link:

https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
In this way you will contribute to improve the safety of this product.

Last revised: 12/2023

For the Summary of Safety and Clinical Performance (SSCP) please see EUDAMED database (https://ec.europa.eu/tools/eudamed/) or contact us at Regulatory.gcc@gc.dental

Undesired effects - Reporting:
If you become aware of any kind of undesired effect, reaction or similar events experienced by use of this product, including those not listed in this instruction for use, please report them directly through the relevant vigilance system, by selecting the proper authority of your country according to the following link:

https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
In this way you will contribute to improve the safety of this product.

Last revised: 12/2023

For the Summary of Safety and Clinical Performance (SSCP) please see EUDAMED database (https://ec.europa.eu/tools/eudamed/) or contact us at Regulatory.gcc@gc.dental

Undesired effects - Reporting:
If you become aware of any kind of undesired effect, reaction or similar events experienced by use of this product, including those not listed in this instruction for use, please report them directly through the relevant vigilance system, by selecting the proper authority of your country according to the following link:

https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
In this way you will contribute to improve the safety of this product.

Last revised: 12/2023

For the Summary of Safety and Clinical Performance (SSCP) please see EUDAMED database (https://ec.europa.eu/tools/eudamed/) or contact us at Regulatory.gcc@gc.dental

Undesired effects - Reporting:
If you become aware of any kind of undesired effect, reaction or similar events experienced by use of this product, including those not listed in this instruction for use, please report them directly through the relevant vigilance system, by selecting the proper authority of your country according to the following link:

https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
In this way you will contribute to improve the safety of this product.

Last revised: 12/2023

For the Summary of Safety and Clinical Performance (SSCP) please see EUDAMED database (https://ec.europa.eu/tools/eudamed/) or contact us at Regulatory.gcc@gc.dental

Undesired effects - Reporting:
If you become aware of any kind of undesired effect, reaction or similar events experienced by use of this product, including those not listed in this instruction for use, please report them directly through the relevant vigilance system, by selecting the proper authority of your country according to the following link:

https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
In this way you will contribute to improve the safety of this product.

Last revised: 12/2023

For the Summary of Safety and Clinical Performance (SSCP) please see EUDAMED database (https://ec.europa.eu/tools/eudamed/) or contact us at Regulatory.gcc@gc.dental

Undesired effects - Reporting:
If you become aware of any kind of undesired effect, reaction or similar events experienced by use of this product, including those not listed in this instruction for use, please report them directly through the relevant vigilance system, by selecting the proper authority of your country according to the following link:

https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
In this way you will contribute to improve the safety of this product.

Last revised: 12/2023

For the Summary of Safety and Clinical Performance (SSCP) please see EUDAMED database (https://ec.europa.eu/tools/eudamed/) or contact us at Regulatory.gcc@gc.dental

Undesired effects - Reporting:
If you become aware of any kind of undesired effect, reaction or similar events experienced by use of this product, including those not listed in this instruction for use, please report them directly through the relevant vigilance system, by selecting the proper authority of your country according to the following link:

https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
In this way you will contribute to improve the safety of this product.

Last revised: 12/2023

For the Summary of Safety and Clinical Performance (SSCP) please see EUDAMED database (https://ec.europa.eu/tools/eudamed/) or contact us at Regulatory.gcc@gc.dental

Undesired effects - Reporting:
If you become aware of any kind of undesired effect, reaction or similar events experienced by use of this product, including those not listed in this instruction for use, please report them directly through the relevant vigilance system, by selecting the proper authority of your country according to the following link:

https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
In this way you will contribute to improve the safety of this product.

Last revised: 12/2023

For the Summary of Safety and Clinical Performance (SSCP) please see EUDAMED database (https://ec.europa.eu/tools/eudamed/) or contact us at Regulatory.gcc@gc.dental

Undesired effects - Reporting:
If you become aware of any kind of undesired effect, reaction or similar events experienced by use of this product, including those not listed in this instruction for use, please report them directly through the relevant vigilance system, by selecting the proper authority of your country according to the following link:

https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
In this way you will contribute to improve the safety of this product.

Last revised: 12/2023

For the Summary of Safety and Clinical Performance (SSCP) please see EUDAMED database (https://ec.europa.eu/tools/eudamed/) or contact us at Regulatory.gcc@gc.dental

Undesired effects - Reporting:
If you become aware of any kind of undesired effect, reaction or similar events experienced by use of this product, including those not listed in this instruction for use, please report them directly through the relevant vigilance system, by selecting the proper authority of your country according to the following link:

https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
In this way you will contribute to improve the safety of this product.

Last revised: 12/2023

For the Summary of Safety and Clinical Performance (SSCP) please see EUDAMED database (https://ec.europa.eu/tools/eudamed/) or contact us at Regulatory.gcc@gc.dental

Undesired effects - Reporting:
If you become aware of any kind of undesired effect, reaction or similar events experienced by use of this product, including those not listed in this instruction for use, please report them directly through the relevant vigilance system, by selecting the proper authority of your country according to the following link:

https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
In this way you will contribute to improve the safety of this product.

Last revised: 12/2023

For the Summary of Safety and Clinical Performance (SSCP) please see EUDAMED database (https://ec.europa.eu/tools/eudamed/) or contact us at Regulatory.gcc@gc.dental

Undesired effects - Reporting:
If you become aware of any kind of undesired effect, reaction or similar events experienced by use of this product, including those not listed in this instruction for use, please report them directly through the relevant vigilance system, by selecting the proper authority of your country according to the following link:

https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
In this way you will contribute to improve the safety of this product.

Last revised: 12/2023

For the Summary of Safety and Clinical Performance (SSCP) please see EUDAMED database (https://ec.europa.eu/tools/eudamed/) or contact us at Regulatory.gcc@gc.dental

Undesired effects - Reporting:
If you become aware of any kind of undesired effect, reaction or similar events experienced by use of this product, including those not listed in this instruction for use, please report them directly through the relevant vigilance system, by selecting the proper authority of your country according to the following link:

https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
In this way you will contribute to improve the safety of this product.

Last revised: 12/2023

For the Summary of Safety and Clinical Performance (SSCP) please see EUDAMED database (https://ec.europa.eu/tools/eudamed/) or contact us at Regulatory.gcc@gc.dental

Undesired effects - Reporting:
If you become aware of any kind of undesired effect, reaction or similar events experienced by use of this product, including those not listed in this instruction for use, please report them directly through the relevant vigilance system, by selecting the proper authority of your country according to the following link:

https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
In this way you will contribute to improve the safety of this product.

Last revised: 12/2023

For the Summary of Safety and Clinical Performance (SSCP) please see EUDAMED database (https://ec.europa.eu/tools/eudamed/) or contact us at Regulatory.gcc@gc.dental

Undesired effects - Reporting:
If you become aware of any kind of undesired effect, reaction or similar events experienced by use of this product, including those not listed in this instruction for use, please report them directly through the relevant vigilance system, by selecting the proper authority of your country according to the following link:

https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
https://ec.europa.eu/growth/sectors/medical-devices/contacts_en
In this way you will contribute to improve the safety of this product.

Last revised: 12

